

**ADDENDUM No. 4 to the TENDER DOCUMENTS for Procurement of Medical Imaging Equipment, no. IOP/5-2017/RD**

**Issued on 17.10.2017.**

- 1. Tender Documents are changed in Technical Specification, Lot 1, 2. Digital angiography system for cardio-vascular diagnostic and interventional procedures Institute for cardiovascular diseases “Dedinje”, line 5.3**

Which reads as follows:

DQE, 70% or less on 0 Lp/mm

It is changed and reads as follows:

DQE, 70% **or more** on 0 Lp/mm

- 2. Tender Documents are changed in Technical Specification, Lot 1, 2. Digital angiography system for cardio-vascular diagnostic and interventional procedures Institute for cardiovascular diseases “Dedinje”, line 8.6**

Which reads as follows:

8.6 Integrated FFR plus OCT or FFR plus IVUS. Integrated in angio patient table in angio-suite. Real time coregistration with angiography image.

It is changed and reads as follows:

**8.6 Integrated FFR, iFR (or equivalent) plus integrated OCT (or equivalent) or Integrated FFR, iFR (or equivalent) plus integrated IVUS on system in angio-suite. Integrated in patient table.**

In order to prepare correct bids, regarding lots in which changes were made, the Bids should be prepared in accordance with the latest amendment, published on the website of the Purchaser.

In all other aspects, Bidding documents for the Procurement of Medical Imaging Equipment, no. IOP/5-2017/RD, remains unchanged.

**Public Procurement Commission**