

AMENDMENT No. 1 TO PROCUREMENT DOCUMENTS

Issued on 11th of October 2022

PROCUREMENT OF EQUIPMENT FOR GENERAL HOSPITAL IN LOZNICA

NO. IOP/61-2021/UHI

In accordance with the Clause 8. Part 1. Bidding Procedures, Section I. Instructions to Bidders, Contents of Bidding Documents, Amendment of Bidding Documents of the Procurement Documents, Public Investment Management Office, No. 11 Nemanjina street, Republic of Serbia, as the Purchaser, hereby notifies all persons concerned for **Procurement of equipment for General Hospital in Loznica, no. IOP/61-2021/UHI**, that there has been an amendment made in the Procurement Documents.

The following provisions of Tender Documents are hereby replaced as follows:

1. Tender Documents are changed in Section VI. Schedule of Requirements, 1. Related Services and Completion Schedule, bullet 1:

- “Delivery period will commence on the date of the advance payment. If the bidder does not require an advance payment implementation period will commence at the entry into force of the Contract Agreement (signing of the contract and submission of the Performance Guarantee) - **120 (one hundred and twenty days) days**”.

It is replaced and reads as follows:

- “Delivery period will commence on the date of the advance payment. If the bidder does not require an advance payment implementation period will commence at the entry into force of the Contract Agreement (signing of the contract and submission of the Performance Guarantee) - **180 (one hundred eighty) days.**”

2. Tender Documents are changed in Technical Specification, Lot 5 – Radiology imaging, ID 2 - Mobile X-ray device:

Line item 2.5 which read as follows:

- Focal spot range min. 65-200

It is replaced and reads as follows:

- Focal spot range min. 68-200

Line item: 2.10 which read as follows:

High frequency generator min. 100 kHz

It is replaced and reads as follows:

High frequency generator min. 60 kHz

Line item: 2.14. which read as follows:

- mAs range min 0.5-360 mAs

It is replaced and reads as follows:

- mAs range min 0.5-320 mAs

Line item: 2.16. which read as follows:

- Two focal spots: max. 0.6 and min. 1.2 mm

It is replaced and reads as follows:

- Two focal spots: max. 0.7 and min. 1.2 mm

Line item: 2.21. which reads as follows:

- Detectors and mobile xray unit are from the same manufacturer

Item 2.21. is deleted.

Line item: 2.25. which reads as follows:

- Removable snap-on anti scatter grid 6:1

It is replaced and reads as follows:

- Removable snap-on anti scatter grid or virtual grid

Line item: 2.26. which reads as follows:

- Detector and grid stored in the bin

It is replaced and reads as follows:

- Detector and grid stored in the bin (only in case when physical grid is offered)

Line item: 2.27. which reads as follows:

- Automatic in bin charging

It is replaced and reads as follows:

- Automatic in bin charging or external charger and additional battery

Line item: 2.33. which reads as follows:

- Image storage capacity min 5.000 images

It is replaced and reads as follows:

- Image storage capacity min 3.500 images

Line item: 2.37. which reads as follows:

- Images export to CD/DVD media for storage

It is replaced and reads as follows:

- Images export to CD/DVD or USB media for storage

3.Tender Documents are changed in Technical Specification, Lot 5 – Radiology equipment, ID 3 – Digital mammography system:

Line item: 3.9. which reads as follows:

- Anode material dual track: Molybdenum (Mo) and Rhodium (Rh)

It is replaced and reads as follows:

- Anode material: Molybdenum (Mo) and Rhodium (Rh) or Tungsten (W)

4.Tender Documents are changed in Technical Specification, Lot 5 – Radiology imaging, ID 9 – Ultrasound device for Neurology:

Line item: 9.39-941. which reads as follows:

9.39	B/W thermal printer
9.40	Probes:
9.41	Linear probe Frequency range: 4 – 13 MHz or wider Field of view (FOV): 40 mm (+/- 5 mm)

Items 9.39 – 9.41 is deleted.

In all other aspects, Tender documents for the **Procurement of equipment for General Hospital in Loznica**", no. IOP/61-2021/UHI, remains unchanged.

This Amendment to Procurement Documents is a constituent part of the **Procurement of equipment for General Hospital in Loznica**", no. IOP/61-2021/UHI, and will be posted on the Purchaser's web site: <http://www.obnova.gov.rs/english/public-procurement>

An updated version of technical specifications is given in the excel file and is a constituent part of this amendment.

In order to submit a responsive bid, bidders are expected to prepare their offers in accordance with this amendment.

Public Procurement Committee

