

**Amendment No. 2 to the TENDER DOCUMENTS for the Procurement of Medical Devices  
for University Children Hospital “Tiršova”, no. IOP/11-2017/RD**

Issued on 30.11.2017.

- 1) Tender Documents are changed in Technical Specification, LOT 2 Cardiology, Position 2.11 Monitor for Cardiology. Position 2.11 is hereby replaced and read as follows:

<b>2,11</b>	<b>Monitor for cardiology</b>
<b>1</b>	Patient vital sign monitor-noninvasive
<b>2</b>	Non invasive ECG monitoring with single channel ECG recording (printable)
<b>3</b>	NIBP - non invasive blood pressure
<b>4</b>	SpO2 - Puls oximetry
<b>5</b>	Power and battery power

- 2) Tender Documents are changed in Technical Specification, LOT 3 General Pediatric departments, Position 3.15 Polysomnography diagnostic device. Position 3.15 is hereby replaced and read as follows:

<b>3,15</b>	<b>Polysomnography diagnostic device</b>
<b>1</b>	Flow
<b>2</b>	Snoring
<b>3</b>	Body position
<b>4</b>	Pulse oxymetry
<b>5</b>	Effort (thoracic and abdominal):Z-Rip inductive technology
<b>6</b>	Total channel number:21 measured + 8 for connection of therapy devices + 6 derived ECG/PTT
<b>7</b>	Battery - capacity:20h of recording
<b>8</b>	Signal resolution/Bit:16Bit
<b>9</b>	Maximum sampling rate:1.000 Hz
<b>10</b>	Maximum archiving rate:200 Hz
<b>11</b>	Interface for connecting electrodes:Intuitive user interface with a human head image
<b>12</b>	Integration with therapy device:Digital
<b>13</b>	PC connection:USB
<b>14</b>	Internal memory archive:SD card 1GB, with compatibility for 4 GB

Updated version of technical specifications are given in excel file and is constituent part of this amendment.

In order to prepare correct bids, regarding lots in which changes were made, the Bids should be prepared in accordance with the latest amendment, published on the website of the Purchaser.

Public Procurement Committee