

**Procurement of Medical Devices for University Children Hospital "Tiršova"**

**IOP/11-2017/RD**

**Clarification no. 14**

**Issued on December 8, 2017**

**LOT 2 - 2.2 MONITOR**

**Question 1:**

Item 9. • Continuous measurement "QT / QTc"

Please change this to "QT / QTc, or Aries measurement, or adequate"

**Answer 1:**

The request is not acceptable

Tender request is monitor with continuous measurement "QT / QTc"

**Question 2:**

Item 16. Thermal printer with three or more channels

Please change this to "Thermal printer with two or more channels".

Two channels are sufficient for a patient monitor, otherwise an ECG device is used.

**Answer 2:**

Accepted

**LOT 5, Item 14 Procedural Guidance Ultrasound**

**Question 3:**

Is it possible to change 12.1"LED touch screen to 12.1" LCD under number 3. on the list?

**Answer 3:**

It is not acceptable.

**Question 4:**

Is it possible to change Linux operating system to Windows operating system under number 17 on the list?

**Answer 4:**

It is not acceptable.

**Question 5:**

Could you define more closely number 10. on the list?

**Answer 5:**

The sensors inside the probe identifies and measure the magnetic field of the magnetized needle. The software calculate the trajectory of the needle and the position where the needle tip will hit your ultrasound beam. You can see that live as a red target box in your ultrasound image (out of plane). When you hit the ultrasound beam the target box become green and you see the reflection of your needle tip inside this box. Then you know that the navigation was successful

**Question 6:**

Is it ISO 13845 certificate for Manufacturer and ISO 9001 for Sevice Company acceptable?

**Answer 6:**

It is acceptable.

Procurement Committee