

Procurement of Medical Devices for Clinical Centre Dr Dragiša Mišović
IOP/6-2017/RD
Clarification no.25
Issued on 02.10.2017.

Q1.

1.

2. Postqualification Requirements (ITB 37.2) – f) Certifications, standards and licences

“Bidder is obliged to submit evidence (licence) that is registered for selling medicines and medical devices on the territory of the Republic of Serbia, issued by Ministry of Health of the Republic of Serbia.

For offered medical devices, it is necessary to submit documentation (licence) of current valid registration in ALIMS -R. Serbia.”

Given the fact that the project is international and open to all eligible participants and it is financed by the European Investment Bank and The Council of Europe Development Bank, these requirements are restrictive and allows possibility to submit an offer only to local companies and specific equipment that was sold so far on the territory of R. Serbia, we ask you to either to eliminate these requests from the tender documents or to accept registration to local authorities mentioned above after the signature of contract as we are aware that this is the local legislation for the medical equipment.

A1

See Addendum no.4, issued on 02.10.2017 on the web site of the Purchaser

Q2


Lot 2 – item 2.2 Automatic steam sterilizer

a) You requested “*Sterilizer has its own steam generator with heaters submerged in water max.power 7.5 kW*”.

Please clarify you accept the sterilizer to gave own steam generator with heaters submerged in water max.power 18 kW.

b) You requested “*Dimensions (HxWxD) max 1550 x 600 x 850 mm*”.

Please clarify you accept dimensions (HxWxD) 1615 x 800 x 1200 mm.



A2

- a) There are no technical conditions at the installation site for the use of a sterilizer with a higher power of the heaters than stated.
- b) The maximum allowed dimensions of the device are given in accordance with the available space for installation and unhindered manipulation.

Q3

Lot 2 – item 2.4.1. Automatic steam sterilizer

You requested “*Dimensions of the sterilizer should be maximum Height x Width x Depth = 1920x1600x1000mm*”.

Please clarify you accept dimensions of the sterilizer (HxWxD) 1920 x 1600 x 1170 mm.

A3

The maximum allowed dimensions of the device are given in accordance with the available space for installation and unhindered manipulation.

Q4

Lot 2 – item 2.4.1. Automatic steam sterilizer

You requested “*Dimensions: max. 1055x1690x1080 mm (WxHxD)*”.

Please clarify you accept dimensions (WxHxD) 1055 x 1700 x 1460 mm

A4

The maximum allowed dimensions of the device are given in accordance with the available space for installation and unhindered manipulation.