

Clarification 2
Procurement of furniture for University Children Hospital Tiršova
IOP/32-2019/UHI

Issued on 8th, November 2019

1. Under Technical specification Medical furniture Line 14. Storage cabinet for storage of flammable liquids under ID 14.15 Requested " Statement of the Bidder that they will submit an Expert Opinion with the signed Contract (copy) of the authorized institution of the Republic of Serbia that the offered type of cabinet for combustible chemicals belongs to class 90 fire resistance, in accordance with EN 14470-1."

Question 1.: Is it acceptable: Statement of the Bidder that they will submit an Expert Opinion of the authorised Institution in the Republic of Serbia that the offered type of cabinet for combustible chemicals belongs to class 90 fire resistance, in accordance with EN 14470-1.? There is no practice that Contract is signed for the Expert Opinion purpose. The authorized institution of the Republic of Serbia- IMS- Institute for materials od Serbia is not signing the Contract to issue Expert opinion.

1. It is acceptable

2. Under Technical specification Medical furniture Line 14. Storage cabinet for storage of flammable liquids under ID 14.14 Requested "CE declaration of cabinets for flammable chemicals as evidence of compliance with the following directives and regulations (copy) 2006/42 / EC 2014/30 / EU 2014/35 / EU EN 14470-1: 2004 EN ISO 121000: 2010 EN ISO 60204/1: 2006"

Question 2. Is it acceptable that Bidder submit statement that following standards are not applicable for offered model: 2014/30 / EU , 2014/35 / EU, EN ISO 60204/1: 2006. ? Offered model is not an electrical equipment and above mentioned standards are for models which are electrical equipment.

2. It is acceptable

Question 3. Please clarify and confirm that there is typing error in ID 14.14. for requested standard 2006/42/EC. Correct standard should be 2006/42/EG.

3. We confirm that it is a typing error. The correct standard is 2006/42/EG

Question 4. Please clarify and confirm that requested Expert Opinion and Electromagnetic Compatibility Regulations, for the model offered cabinet and ventilator (ID 14.15 and ID 14.16) should be issued in the name of Producer Authorised Distributor for territory of the Republic of Serbia. Authorised Distributors for Import in the Republic of Serbia are responsible to obtain requested documents in their name not in the name of the bidder. "

4. It is acceptable.

"Technical specification - Medical furniture ID 15. Requested: "Medical X ray negatoscope double, dimensions 600 mm x 450 mm + 5%, Brightness regulation 10%-100%" Question: Is it acceptable X ray negatoscope dimensions 740mm x 470 mm + 5% with uniform brightness over viewing fight surface? Requested dimension and brightness are not appropriate for two radiology films for simultaneous viewing. Brightness regulation is important only for mammography film reading. "

5. It is acceptable.

Public Procurement Commission