

Procurement of Furniture for University children hospital Tiršova

IOP/32-2019/UHI

Clarification no. 1

Issued on October 28, 2019

Question 1:

Technical Specification- Medical Furniture, Line Item No.1 Stretcher without fences, ID 1 position 1. Requested:"Two-piece plate from multi-layer laminate, permeable to X rays". Is it acceptable to offer better solution with more than two-piece plate, eg. four -piece plate? The solution is better for operator manipulation with stretcher.

Answer 1:

It is acceptable.

Question 2:

Technical Specification- Medical Furniture, Line Item No.2 Stretcher with fences, ID 2 position 1. Requested:"Two-piece plate from multi-layer laminate, permeable to X rays". Is it acceptable to offer better solution with more than two-piece plate, eg. four -piece plate? The solution is better for operator manipulation with stretcher.

Answer 2:

It is acceptable.

Question 3:

Technical Specification-Medical Furniture, Line No.3. Hydraulic therapy table, ID 3 position 5. Requested: "Height adjustment is via a circular switch that is operated by a foot on either side". Is it acceptable to offer Height adjustment is via a foot pedal. This solution is more appropriate for hydraulic therapy table.

Answer 3:

It is acceptable.

Question 4

Requested under e) "The Bidder is obliged to apply the quality management system in accordance with the requirements of the standards, as follows: ISO 9001, ISO 14001, ISO 45001 and FSC COC." Question: Is it acceptable for Bidder ISO 9001, ISO 14001, ISO 45001 and FSC COC for Manufacturer of wood panels?

Answer 4:

It is acceptable to provide FSC COC for Manufacturer of wood panel, because this standard is appropriate for manufacturers of such products.

Requirement for ISO 9001, ISO 14001 and ISO 45001 for Bidder remains the same.