

**Procurement of medical equipment for Clinical-hospital center  
„Dr Dragiša Mišović - Dedinje“  
IOP/59-2021/UHI**

**Clarification No.1**

**Issued on 18<sup>th</sup> of February 2022**

**Question 1:**

"Question: Please revise the request under ID 16 to be "Minimal range of flow trigger: 0,2 - 10 L/min" because this is clinically realistic and sufficient range of use that all anaesthesia systems satisfy, so more bidders could prepare acceptable offer ?

Explanation: The flow trigger is used for detection of spontaneous breathing efforts of the patient. Its clinical significance is in lower values of range, since this means better sensitivity of spontaneous breath detection. In clinical practice flow trigger of high volume, such is up to 15 L/min is never used, because it would mean the anaesthesia machine should measure flow difference from even 15 L/min between inspiration and expiration that can detect a spontaneous breath effort of the patient. Current requirement do not represent realistic way how the flow trigger is used."

Answer 1:

Request is acceptable. Purchaser will publish Amendment no. 1. to Procurement Documents in accordance with the requested.

Public Procurement Committee

