

**PROCUREMENT OF EQUIPMENT FOR BIOSENSE INSTITUTE IN  
NOVI SAD, SERBIA  
(PROCUREMENT NO. IOP/58-2021/RD)**

Clarification no. 3

Issued on May 13, 2022

Regarding the list of questions that the Purchaser, Public Investment Management Office Belgrade, No. 11 Nemanjina street, has received from the potential bidders, concerning the procurement procedure: Equipment for Biosense Institute in Novi Sad, Serbia no. IOP/58-2021/RD, we give you the following answers:

**Question 1:**

Lot 7 – Biosystems

"LOT 7, ID 7.6 Material printer: Detailed Technical Specification describes Materials Cartridge only.

Please clarify if we should offer complete Material printer or only Cartridge?

If complete printer is requested, please provide required specifications for the printer."

**Answer 1:**

The complete system is requested. Below you can find additional technical specs:

- Base printer unit with drop watcher and fiducial camera.
- Embedded PC with drop manager software and included application software (pre-loaded patterned templates, pattern preview, editors for pattern, bitmap files supported).
- Printing area at least 200 x 250 mm.
- Printing substrates thickness up to 25 mm.
- Vacuum substrate holder with adjustable temperature ranging from ambient to 60°C.

Please find the updated Excel file and the Amendment No. 3 to Bidding Documents reflecting this attached.

Public Procurement Committee

  
\_\_\_\_\_  
Jelena Simić