



## No.1

## Procurement of MRI Scanner

Република Србија  
Канцеларија за управљање  
јавним улагањима

404-02-11/2017-07

Amendment number	Reference in Tender documents	Original text	Amendment (new text)
1.	Technical Specifications 4.7.	Separate, dedicated multi-channel coil for the imaging of pediatric (babies) head and neck including neuro-vascular examination with minimum of 16 coil elements for basic and advanced head studies capable for simultaneous acquisition with other coils.	Separate, dedicated multi-channel coil for the imaging of pediatric (babies) head and neck including neuro-vascular examination with minimum of 8 coil elements for basic and advanced head studies capable for simultaneous acquisition with other coils -
2.	Technical Specifications 4.8.	Separate, dedicated multi-channel coil for the imaging of small Field of Views and small structures near the surface.	Separate, dedicated multi-channel coil or two single channel coils for the imaging of small Field of Views and small structures near the surface.
3.	Technical Specifications 10.6.	Number of coil elements for spine coil: 12 - 0 point 13 to 25 - 1 point 26 to 35 - 2 points 36 or higher - 3 points	Number of coil elements for spine examination (with spine coil): 12 - 0 point 13 to 25 - 1 point 26 to 35 - 2 points 36 or higher - 3 points
4.	Technical Specifications 10.11.	Artifact free abdominal imaging without gating or breathold using Radial imaging technique: NO - 0 point YES - 5 points	Artifact reducing technique for abdominal imaging without gating or breathold using Radial imaging technique: NO - 0 point YES - 5 points

The Public Procurement Commission