

ADDENDUM No. 2 to the TENDER DOCUMENTS for Procurement of Medical Devices for Clinical Centre Dr Dragiša Mišović, no. IOP/6-2017/RD

Issued on 25.09.2017.

- 1) Tender Documents are changed in Technical Specification, LOT 3 Hospital furniture, Position 3.1 Hospital bed for ward. Position 3.1 is hereby replaced and read as follows:

1	The basic construction of the bed on two telescopic columns with adjustable height
	External dimensions
2	Length: min.2250 mm
3	Width: max. 965 mm
4	Electrical height adjustment of the bed without a mattress, min. range 390 - 770 mm
5	Permissible load min 250 kg
6	Bed pallet extension min. 200 mm
7	Bed frame made of steel and coated with antibacterial paint resistant to disinfectants and detergents
8	Maximum four segmented bed pallet, made of HPL, the whole surface of the bed must be permeable for the X rays
9	Side bumpers placed on four edges of the bed
10	The bed must have tunnel cassette for X rays, below the backrest part
11	The bed must be compatible to work with C arm
12	Min. 4 hooks for drainage and urine bags on each side of the bed
13	Side rails made of metal covered with antibacterial paint resistant to disinfectants and cleaning agents.
14	Headboard and footboard made of cast polypropylene, steel or aluminum, easy disassembly by pulling up.
15	Wheels diameter min. 150 mm
16	Central brake blocks all 4 wheels
17	The bed must be equipped with 4 places for drip bottle holder
18	Double, simultaneous auto-regression function – for back rest min 120 mm, for thigh rest min. 45 mm
19	Double auto-contour function –inclination angle of backrest min. 70°, thigh rest min 40°
20	The possibility of electrical settings for the back of the segment from 0 to 70 °, with inclination indicator
21	Trendelenburg / Reverse Trendelenburg: min. + 15 ° / -15 °
22	Electrical adjustment of thigh segment: min. 0 -40°
23	Adjustable shins rest segment - mechanical
24	Patient control panel (wired remote control)
25	Nurses control panel with all the controls placed at footboard, locks the control functions of the patient control panel. Low position, anti-shock position, reverse Trendelenburg, cardiac chair, electric CPR, examination position
26	Battery for electric function of the bed when not online
27	Degree of protection IP-66
	Additional equipment
28	Passive anti-decubitus mattress adequate to the dimensions of the bed pallet, with antibacterial cover, waterproof and steam-permeable, thickness min. 14cm,.
29	Drip bottle holder with 4 hooks, 1 piece

30	Lifting pole with hand holder, 1 piece
----	--

Updated version of technical specification for this position is given in excel file and is constituent part of this amendment.

In order to prepare correct bids, regarding lots in which changes were made, the Bids should be prepared in accordance with the latest amendment, published on the website of the Purchaser.

Public Procurement Commission


