

**ADDENDUM No. 3 to the TENDER DOCUMENTS for Procurement of Medical Devices for Clinical Centre Dr Dragiša Mišović, no. IOP/6-2017/RD**

**Issued on 28.09.2017.**

- 1) Tender Documents are changed in Technical Specification, LOT 4 Auxiliary surgical devices:

Following positions are deleted:

**4.17.1-Neonatal Resuscitation (Resuscitation table for newborns),**

**4.19 - Mobile bathroom**

**4.20 - Coasters medical-grade silicone for operating tables – set**

- 2) Tender Documents are changed in Technical Specification, LOT 5 Anesthesia, position **5.1 Anesthesia system with monitoring:**

Line 30 reads as follows:

**It has trend of 72 h**

It is replaced and reads as follows:

**It has trend of 24 h**

- 3) Tender Documents are changed in Technical Specification, LOT 5 Anesthesia, **LOT 5 Anesthesia, Position 5.6 – Ventilator with standard capnography:**

Line 23 reads as follows:

PEEP

It is replaced and reads as follows:

**PEEP:1- 50 cm H2O**

Line 24 reads as follows:

Insp3T: 1- 50 cm H2O;

It is replaced and reads as follows:

**Insp T: 0,5-10 sec**

Updated version of technical specifications are given in excel file and is constituent part of this amendment.

In order to prepare correct bids, regarding lots in which changes were made, the Bids should be prepared in accordance with the latest amendment, published on the website of the Purchaser.