

AMENDMENT No. 3 TO PROCUREMENT DOCUMENTS

Issued on 01th of November 2022

PROCUREMENT OF EQUIPMENT FOR GENERAL HOSPITAL IN LOZNICA

NO. IOP/61-2021/UHI

In accordance with the Clause 8. Part 1. Bidding Procedures, Section I. Instructions to Bidders, Contents of Bidding Documents, Amendment of Bidding Documents of the Procurement Documents, Public Investment Management Office, No. 11 Nemanjina street, Republic of Serbia, as the Purchaser, hereby notifies all persons concerned for **Procurement of equipment for General Hospital in Loznica, no. IOP/61-2021/UHI**, that there has been an amendment made in the Procurement Documents.

The following provisions of Tender Documents are hereby replaced as follows:

1. Tender Documents are changed in Technical Specification, Lot 5 –Radiology Imaging, Item 6 - Ultrasound system with 3 probes:

- ID 6.17 which read as follows:

Max Read Zoom to 20x: baseline shift on “freeze” or recall.

ID 6.17. Max Read Zoom to 20x: baseline shift on “freeze” or recall - is deleted.

- ID 6.19. which read as follows:

The ability to import CT/MRI images for comparison on a split screen with active ultrasound image

ID 6.19. The ability to import CT/MRI images for comparison on a split screen with active ultrasound image – is deleted

- ID 6.24 which read as follows:

Linear with 4 configurable buttons.

ID 6.24. Linear with 4 configurable buttons - is deleted.

- ID 6.32 which read as follows:

Curved anatomical M-Mode.

ID 6.32. Curved anatomical M-Mode - is deleted.

- ID 6.37 which read as follows:

Ability to display ultrasound image at the same time on monitor and on the touch panel

ID 6.37. Ability to display ultrasound image at the same time on monitor and on the touch panel - is deleted.

2. Tender Documents are changed in Technical Specification, Lot 5 –Radiology Imaging, Item 9 - Ultrasound device for Neurology:

- **ID 9.17. which read as follows:**

HD write zoom - read zoom min. 50X

ID 9.17. HD write zoom - read zoom min. 50X -is deleted

- **ID 9.18. which read as follows:**

Voice comments recording

ID 9.18. Voice comments recording - is deleted

3. Tender Documents are changed in Technical Specification, Lot 1 –Endoscopic Suergery and Diagnostics, Item 5 – Videoendoscopic tower with gastroscope and conoscope:

- **ID 5.21. which read as follows:**

Compatible with chromatographic observation of tissues described under 1.11, 1.12 and 1.13.

It is replaced and reads as follows:

ID 5.21. Compatible with chromatographic observation of tissues described under 5.7, 5.8 and 5.9

- **ID 5.30. which read as follows:**

ID 5.30. - Compatible with chromatographic observation of tissues described under 1.11, 1.12 and 1.13

It is replaced and reads as follows:

ID 5.30. Compatible with chromatographic observation of tissues described under 5.7, 5.8 and 5.9

4. Tender Documents are changed in Technical Specification, Lot 10 –Sterilization, after Item 23 – Steam sterilizer for steam disinfection of infectious waste, Item no. 24. Water preparation system set is added. It will be amended in the Amendment no. 3.

5. Tender Documents are changed in Technical Specification, Lot 5 –Radiology Imaging, Item 1 - Ceiling mounted x-ray device:

- **ID 1.34 which reads as follows:**

ID 1.34. - Vertical motion control by foot switch.

It is replaced and reads as follows:

ID 1.34. - Vertical motion control by foot switch or via buttons directly on BWS or via wireless remote control.

6. Tender Documents are changed in Technical Specification, Lot 5 –Radiology Imaging, Item 1 - Ceiling mounted x-ray device:

- ID 1.7. Which reads as follows:

ID 1.7. Automatic detector charging in bucky

ID 1.7. Automatic detector charging in bucky – is deleted

- ID 1.57. which reads as follows:

ID 1.57. In-bin charging, the detectors automatically charge when they are in the bucky housing

ID 1. 57. In-bin charging, the detectors automatically charge when they are in the bucky housing – is deleted

7. Tender Documents are changed in Technical Specification, Lot 9 –Hospital Furniture, Item 33 - Upholstered chair without backrest - for operating room:

- ID 33.1. which reads as follows:

ID 33.1. Upholstered seat with antistatic cover, height adjustment from 560 to 790mm by means of gas spring, with foot ring release, back rest with antistatic cover

It is replaced and reads as follows:

ID 33.1 Upholstered seat with antistatic cover, height adjustment from 560 to 790 mm by means of gas spring, with foot ring release, with antistatic cover.

8. Tender Documents are changed in Technical Specification, Lot 6 –Laboratory Patology, Item 20 – Ultra Low Temperature Freezer – Double Security:

- ID 20.10. which reads as follows:

ID 20.10. Castors

ID 20.10. Castors – is deleted

9. Tender Documents are changed in Technical Specification, Lot 6 –Laboratory Patology, Item 17 – Fridge for storing medicine

- ID 17.10. which reads as follows:

ID 17.10. Self closing door

ID 17.10. Self closing door – is deleted

10. Tender Documents are changed in Technical Specification, Lot 7 –Mobile Diagnostics equipment, Item 21 – EMNG

- ID 21.50. which reads as follows:

ID 21.50. Upper frequency (-3dB) min 0.2 – 5000 Hz

It is replaced and reads as follows:

ID 21.50 - High frequency (-3dB) min. 30-20000Hz

- ID 21.63. which reads as follows:

21.63. Ultrasound display with a linear probe with a frequency of 5 - 15 MHz for examination of the carpal tunnel”

It is replaced and reads as follows:

21.63. Upgrade possibility with ultrasound display with a linear probe with a frequency of 5 - 15 MHz for examination of the carpal tunnel

11. Tender Documents are changed in Technical Specification, Lot 9 –Hospital furniture, Item 19 – Bedside cabinet with overbed table

- ID 19.7. which reads as follows:

ID 19.7. Sizes: 450 x 4540 x 750 H (Except Castors) $\pm 5\%$.

It is replaced and reads as follows:

ID 19.7. Sizes: 450 x 450 x 750 H (Except Castors) $\pm 5\%$.

12. Tender Documents are changed in Technical Specification, Lot 6 –Laboratory Patology, Item 13 –Automatic glucoanalyzer

- ID 13.6. which reads as follows:

ID 13.6. Barcode Reader: external PS/2

It is replaced and reads as follows:

ID 13.6. Barcode Reader

13. Tender Documents are changed in Section VI Schedule of Requirements, 3. General Technical Requirements, point 2. Equivalency of Standards and Codes, point 2.2:

- The equipment offered should be manufactured in compliance with Quality Standard ISO 9001 certification for Manufacturer(s).

Tender Documents, Section VI Schedule of Requirements, 3. General Technical Requirements, point 2. Equivalency of Standards and Codes, point 2.2 - **will be deleted**.

In all other aspects, Tender documents for the **Procurement of equipment for General Hospital in Loznica**, no. IOP/61-2021/UHI, remains unchanged.

This Amendment to Procurement Documents is a constituent part of the **Procurement of equipment for General Hospital in Loznica**, no. IOP/61-2021/UHI, and will be posted on the Purchaser's web site: <http://www.obnova.gov.rs/english/public-procurement>

An updated version of technical specifications is given in the excel file and is a constituent part of this amendment.

In order to submit a responsive bid, bidders are expected to prepare their offers in accordance with this amendment.

Public Procurement Committee

