

AMENDMENT No. 1 TO PROCUREMENT DOCUMENTS

Issued on 30th of June 2021

PROCUREMENT OF EQUIPMENT FOR NEW BUILDING OF INSTITUTE FOR CARDIOVASCULAR DISEASE – DEDINJE 2, BELGRADE

NO. IOP/51-2021/UHI

In accordance with the Clause 8. Part 1. Bidding Procedures, Section I. Instructions to Bidders, Contents of Bidding Documents, Amendment of Bidding Documents of the Procurement Documents, Public Investment Management Office, No. 11 Nemanjina street, Republic of Serbia, as the Purchaser, hereby notifies all persons concerned for **Procurement of equipment for new building of Institute for cardiovascular disease – Dedinje 2, Belgrade**, No. IOP/51-2021/UHI, that there has been an amendment made in the Procurement Documents.

The following provisions of Tender Documents are hereby replaced as follows:

1. **Tender Documents are changed in Technical Specification, Lot 4, Hybrid operating room, Line item number 1- Digital angiography system for hybrid operating room, ID 37, which reads as follows:**

- DQE: min. 79 % at 0 lp/mm (RQA5)

It is replaced and reads as follows:

DQE: min. 77% at 0 lp/mm (RQA5)

2. **Tender Documents are changed in Technical Specification, Lot 4, Hybrid operating room, Line item number 1- Digital angiography system for hybrid operating room, ID 41, which reads as follows:**

- Signal to electronic noise ratio min. 25dB (RQA 5)

It is replaced and reads as follows:

Signal to electronic noise ratio min. 10dB (RQA 5)

In all other aspects, Tender documents for the **Procurement of equipment for new building of Institute for cardiovascular disease – Dedinje 2, Belgrade**, No. IOP/51-2021/UHI, remains unchanged.

This Amendment to Procurement Documents is a constituent part of the **Procurement of equipment for new building of Institute for cardiovascular disease – Dedinje 2, Belgrade**, No. IOP/51-2021/UHI, and will be posted on the Purchaser's web site: <http://www.obnova.gov.rs/english/public-procurement>

Updated version of technical specifications are given in excel file and is constituent part of this amendment.

In order to submit a responsive bid, bidders are expected to prepare their offers in accordance with this amendment.

Public Procurement Committee

Marija Stanković