## INFORMATION AMENDMENT No. 2 TO PROCUREMENT DOCUMENTS Issued on 25<sup>th</sup> of July 2018

for Procurement of Medical devices for the Clinic for Neurosurgery-Clinical Centre of Serbia, No. IOP/12-2018/RD

In accordance with the Clause 8. Part 1. Bidding Procedures, Section I. Instructions to Bidders, Contents of Bidding Documents, Amendment of Bidding Documents of the Procurement Documents, Public Investment Management Office, No. 11 Nemanjina street, Republic of Serbia, as the Purchaser, hereby notifies all persons concerned for Procurement of Medical devices for the Clinic for Neurosurgery-Clinical Centre of Serbia, No. IOP/12-2018/RD, that there has been an amendment made in the Technical specification, lot 6 – Ultrasonic Aspiration:

The new Item 6.3 was added with the following technical specification:

6.3	Ultrasonic Surgical Aspiration system for spine and brain surgery	Qty.
1	Console with integrated peristaltic and aspiration pump	
2	Foot switch for activation	
3	Three operating frequences: 25 kHz, 35 kHz and 55 kHz	
4	Integrated holder for irrigation bottle and aspiration container	
5	Protective filter for preventing the penetration of the aspirate into the interior of the machine	
6	Handpiece for Neurosurgery with external aspiration: working lenght min. 120 mm, sonotrode tip inner diameter 0,9 mm	
7	Handpiece for Neurosurgery with integrated aspiration: working lenght min. 35 mm, sonotrode tip inner diameter 1,4 mm	- 1 - 1
8	Handpiece for Neurosurgery with integrated aspiration: working lenght min. 105 mm, sonotrode tip inner diameter 1,7 mm	,
9	Handpiece for spinal surgery: sonotrode with 40 mm and 100 mm working lenght, rasp area min. 4 mm2	

Updated version of technical specifications are given in excel file and is constituent part of this amendment.

In order to prepare correct bids, regarding lots in which changes were made, the Bids should be prepared in accordance with the latest amendment, published on the website of the Purchaser.

In all other aspects, Procurement Document for the Procurement of Medical devices for the Clinic for Neurosurgery-Clinical Centre of Serbia, No. IOP/12-2018/RD, remains unchanged.

Procurement Commission

Jelena Simić