

Procurement of Medical equipment for Emergency Centre-Clinical Centre of Serbia IOP/23-2018/RD

Issued on March 4, 2019

Question 1

- In the Table of Technical Specifications, Item 1.13 Mobile radiography digital system with wireless FPD (Flat Panel Detector), Line 27 Detector weight: max 3 kg is required.
- Depending on detector thickness and technology, weight of detector is usually a bit more than 3 kg, so in order to get more offers without any decrease in quality please allow bidders to offer detector with weight of 3.3 kg, which is only 10% more than requested 3 kg

Answer 1

It is acceptable.

Question 2

- “In the section 4 – Quality Control Standards – it is requested that all the manufacturers hold the ISO 9001 certificate.
- Please bear in mind that for medical device manufacturers the prevailing certification is ISO 13485, which pertains to production of medical devices. Therefore we suggest that you phrase this requirement in the following manner: “ISO 9001 certification, or ISO 13485 certification (Quality Management System ”.

Answer 2

ISO 13485 is acceptable instead of ISO 9001 certification

Question 3

- In Bidding Documents in Price Schedule template it is envisaged that the bidder gives the price on parity DAP, according to the actual Incoterms. Also, in Bid Submission Form (page 40 of 86), the Total price of Bid should be given on parity DAP (insured and delivered on site, excluding VAT and Customs Duties on import). In Section II. Bidding data Sheet, ITB 14.6 , it is stated that customs duties on import and other related costs are specially declared.
- Since above conditions/incoterms and templates are only applicable in case of foreign bidders and not for bidders from Republic of Serbia (Customs law and VAT law are not taken into consideration because the price of Bid must include import costs, customs duty and 20% VAT for turnover on the territory of Republic of Serbia), we kindly ask for instructions how bidders from Republic of Serbia should give the price of Bid. We suggest parity DDP with indication that VAT will be calculated according to the VAT law of Republic of Serbia. In terms of that we kindly ask for adjustment of all related templates and Tender forms which include pricing.

We kindly ask you to confirm if this project is exempted from customs and VAT and in accordance to that provide instructions how local bidders can be exempted from customs and VAT for turnover within Republic of Serbia.

Answer 3

Project is exempted from customs and VAT. After contracting successful bidders will be provided with instructions.

Question 4

- In Section II. Bidding data Sheet, ITB 14.6 , it is stated that the Purchaser is allowed to accept the offer of goods already imported only in case of having a proof, submitted by the Bidder, of possibility to conduct the revision of the clearance procedure.
- Please advise which document should be presented as a proof stated above.

Answer 4

This provision is not applicable.

Question 5

- In Section VIII. Special conditions of contract, GCC 12.1. are specified documents to be furnished by Supplier before arrival of the goods, such as: copy of invoice, certificate of origin, estimated time and place of entry in Republic of Serbia.
- We assume that these documents are meant only for foreign suppliers and not for suppliers from Serbia and we kindly ask Purchaser to confirm our understanding.

Answer 5

This provision is applicable for foreign bidders.