

**EQUIPMENT FOR BIOSENSE INSTITUTE IN NOVI SAD, SERBIA  
(PROCUREMENT NO. IOP/22-2018/RD)**

Clarification no. 7

Issued on February 8, 2019

Regarding the list of questions that the Purchaser, Public Investment Management Office Belgrade, No. 11 Nemanjina street, have received from the potential bidders, concerning the procurement procedure: Equipment for Biosense Institute in Novi Sad, Serbia no. IOP/22-2018/RD, we give you the following answers:

**Question 1:**


REQUESTED: Optical System: Parallel-optics type (zooming type)

QUESTION: Is it possible to offer the microscope with Greenough optical system

**Answer 1:**

Microscope with Greenough optical system is acceptable.

Public Procurement Committee



Milica Veljković