

**PROCUREMENT OF MEDICAL EQUIPMENT FOR CLINICAL-HOSPITAL
CENTRE "DR DRAGIŠA MIŠOVIĆ-DEDINJE
(PROCUREMENT NO. IOP/59-2022/UHI)**

Clarification no. 2

Issued on February 28, 2022

Regarding the list of questions that the Employer, Public Investment Management Office Belgrade, No. 11 Nemanjina street, have received from the potential bidders, concerning the procurement procedure: Procurement of medical equipment for Clinical-hospital centre "Dr Dragiša Mišović-Dedinje", no. IOP/59-2021/UHI, we give you the following answers:

Question 1:

1. LOT 1 Surgery

- 1.7.1, item 7 Lithotripsy Probe, Reusable, 1.50 mm (4.5 Fr), length 564 mm, 6 pc Please, clarify if there was a technical error as the specification demands 6 pc instead of 1 pc. This is reusable probe and there is no need for purchasing more than 1 pc.

Yes, it is a technical error and Purchaser will amend it in Amendment.

- 1.8.1, item 43 and 44 Is it acceptable for Purchase to offer items with slightly different dimensions: Field of view: is it acceptable to offer item with 88 degrees, as in the specification is stated 90 degrees Working length: is it acceptable to offer 316,4 mm, as in the specification is stated 316 mm?

Yes, it is also acceptable.

- 1.8.1, item 73 Dissection forceps, Maryland, length of jaws: 20 mm, 5x330mm, handle monopolar Is it acceptable to offer the instrument with length 21 mm?

Yes, it is also acceptable –

- 1.4, item 70. The requirement in the technical specification states that it is necessary to offer a 180 °/180 ° field of view monitor. By default, monitors have a field of view of 178 ° / 178 °, so please answer whether it is acceptable to offer a monitor with a field of view of 178 ° / 178 °?

Yes, it is a technical error and Purchaser will amend it in Amendment.

- 1.1.1 item 1 - Round bowl 60x30mm, 0,07l.

Is it acceptable to offer Round bowl 61x30mm, 0,07l ?

Yes, it is also acceptable.

- 1.1.1 item 2 - Round bowl 150x65mm, 0,07l Is it acceptable to offer Round bowl 147x65mm, 0,07l?

Yes, it is also acceptable.

- 1.1.2 item 1 - Round bowl 115x50mm, 0,35l Is it acceptable to offer Round bowl 116x50mm, 0,35l ?

Yes, it is also acceptable.

- 1.1.2 item 2 - Round bowl 150x50mm, 0,70l Is it acceptable to offer Round bowl 147x65, 0,70l ?

Yes, it is also acceptable.

- 1.1.2 item 3 - Gosset abdominal retractor 150mm, lateral valves 65x30mm and 65x35mm, with central valve 60x55mm.

Is it acceptable to offer Gosset abdominal retractor 150mm, lateral valves 63x29mm and 63x36mm, with central valve 59x57mm?

Yes, it is also acceptable.

- 1.1.2 item 6 - Roux retractor double-ended 30x30mm/35x40mm, 17cm.

Is it acceptable to offer Roux retractor double-ended 31x30mm/36x42mm, 17cm?

Yes, it is also acceptable.

- 1.1.4 item 1 - Roux retractor, double-ended, 30x30mm/35x40mm, 17cm.

Is it acceptable to offer Roux retractor, double-ended, 31x30mm/36x42mm, 17cm?

Yes, it is also acceptable.

- 1.1.4 item 2 - Farabeuf retractor set, 25x15mm/30x15mm, 20x15mm/25x56mm, 15,5cm.

Is it acceptable to offer Farabeuf retractor set, 25x16mm/28x16mm, 21x16mm/24x16mm, 15,5cm?

Yes, it is also acceptable.

- 1.1.4 item 36 - Bowl 130x55mm, 0,45l Is it acceptable to offer Bowl 128x55mm, 0,45l?

Yes, it is also acceptable.

- 1.1.4 item 37 - Bowl 165x75mm, 1,1l Is it acceptable to offer Bowl 167x75mm, 1,1l?

Yes, it is also acceptable.

- 1.1.6. item 19 - Instrument insert Claw forceps, 2x3 teeth, diam.10mm, jaw length 30mm, shaft length 330mm.

Is it acceptable to offer Instrument insert Claw forceps, 2x3 teeth, diam.10mm, jaw length 29mm, shaft length 330mm ?

Yes, it is also acceptable.

- 1.1.8. item 20 - De Bakey needle holder with tungsten carbide inserts, 0,5mm, 15cm, gold handles.

Is it acceptable to offer De Bakey needle holder with tungsten carbide inserts, 0,4mm, 15cm, gold handles?

Yes, it is also acceptable.

- 1.1.8. item 29 - Round bowl 115x50mm, 0,35l,

Is it acceptable to offer Round bowl 116x50mm, 0,35l?

Yes, it is also acceptable.

- 1.1.8 item 30 - Round bowl 145x65mm, 0,70l.

Is it acceptable to offer Round bowl 147x65mm, 0,70l?

Yes, it is also acceptable.

- 1.1.8 item 31 - Perforated stainless-steel tray 550x250x70mm.

Is it acceptable to offer Perforated stainless-steel tray 540x253x67mm?

Yes, it is also acceptable.

- 1.1.9 item 29 - Perforated stainless-steel tray 250x250x70mm.

Is it acceptable to offer Perforated stainless-steel tray 253x243x67mm.

Yes, it is also acceptable.

Question 2:

LOT 1 - Surgery

- 1.1. - Surgical instruments.

Is it acceptable to offer instruments with ± 10 mm deviation in dimensions?

Yes, it is also acceptable.

- 1.1.1 item 22 - Dissecting scissors with welded tungsten carbide blades, curved, 18cm, with stable 3-point positions, ergonomically designed scissors rings with V-shaped handle.

Is it acceptable to offer Dissecting scissors with welded tungsten carbide blades, curved, 18cm, ergonomically designed?

Yes, it is also acceptable.

- 1.1.1 item 33 - Scalpel blade remover to remove the blades from handle without the risk of injury, 14cm. Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

Yes, it is also acceptable.

-1.1.2 item 3 - Gosset abdominal retractor 150mm, lateral valves 65x30mm and 65x35mm, with central valve 60x55mm.

Is it acceptable to offer this instrument with 2,5 cm difference in length?

Yes, it is also acceptable.

- 1.1.2 item 6 - Roux retractor double-ended 30x30mm/35x40mm, 17cm.

Is it acceptable to offer this instrument with 10-15 mm deviations of dimensions?

Yes, it is also acceptable.

- 1.1.2 item 20 - Potts-Smith dissecting forceps with tungsten carbide inserts, with dissector end, 0,4mm, 24,5cm, gold handle.

Is it acceptable to offer dissecting forceps with tungsten carbide inserts, length 200 mm, ring shaped dissector end diameter 2,5 mm, gold handle.

Yes, it is also acceptable.

- 1.1.2 item 24 - Scalpel blade remover to remove the blades from handle without the risk of injury, 14cm.

Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

Yes, it is also acceptable.

- 1.1.4 item 9 - Potts-Smith dissecting forceps with tungsten carbide inserts, 0,4mm serration, 18,5cm, with dissector end, golden handles.

Is it acceptable to offer dissecting forceps with tungsten carbide inserts, length 180 mm, ring shaped dissector end diameter 2,5 mm, gold handle?

Yes, it is also acceptable

- 1.1.4 item 15 - Scalpel blade remover to remove the blades from handles without risk of injury, 14cm.

Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

Yes, it is also acceptable.

- 1.1.4 item 18 - Scalpel handle angled, no. 3L, 20,5cm.

Is it acceptable to offer this instrument with 2 cm deviations of length?

Yes, it is also acceptable.

- 1.1.4 item 38 - Stille-Giertz rib sheares, 27cm. Is it acceptable to offer 2 cm shorter sheares?

Yes, it is also acceptable.

- 1.1.4 item 43 - Stitch helper (retractor for wound closing), 200mm. Is it acceptable to offer retractor for wound closing with 1,5 cm deviation in length?

Yes, it is also acceptable.

- 1.1.4 item 53 - Gallaher bone file 17cm. Is it acceptable to offer instrument that is 2 cm longer?

Yes, it is also acceptable.

- 1.1.4 item 54 - Robb intima spatula (3mm) and dilator(1,5mm), 20,5cm. Is it acceptable to offer spatula without dilator?

Yes, it is also acceptable.

- 1.1.5 item 10 - Micro forceps curved, tip 0,8mm, 21cm, titanium with ceramic coating. Is it acceptable to offer forceps with straight forceps with 1 mm tip?

Yes, it is also acceptable.

- 1.1.5 item 18 - Mixer dissecting forceps, angled 90°, diam.10mm, jaw length 20,5mm, shaft length 330mm, handle without ratchet with connector, complete instrument. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 19 - Instrument insert Claw forceps, 2x3 teeth,diam.10mm, jaw length 30mm, shaft length 330mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 20 - Tube shaft length 330mm, for instrument insert diam. 10Mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 22 - Instrument insert grasping forceps, corrugated, atraumatic, diam.5mm, jaw length 21,6mm, shaft length 330mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

-1.1.5 item 23 - Tube shaft length 330mm, for instrument insert diam. 5Mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

Question 3:

- 1.1.5 item 25 - Instrument insert grasping forceps, corrugated, atraumatic, diam.5mm, jaw length 26,6mm, shaft length 330mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 26 - Tube shaft length 330mm, for instrument insert diam. 5mm Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 28 - Instrument insert Dorsey (Johann) grasping forceps fenestrated, diam.5mm, jaw length 20mm, shaft length 330mm. Is it acceptable to offer instrument 310 mm long with jaw length 35 mm?

Yes, it is also acceptable.

- 1.1.5 item 29 - Tube shaft length 330mm, for instrument insert diam. 5Mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 31 - Instrument insert Clinch grasping forceps, fenestrated, atraumatic, diam.5mm, jaw length 20mm, shaft length 330mm. Is it acceptable to offer instrument 310 mm long with jaw length 25 mm?

Yes, it is also acceptable.

- 1.1.5 item 32 - Tube shaft length 330mm, for instrument insert diam. 5Mm. Is it acceptable to offer instrument 310 mm long?

Yes, it is also acceptable.

- 1.1.5 item 39 - Handle with 2-way stopcock, sliding valve and tube connectors. This item is the part of item 40 and they have the same catalogue number. Is that acceptable for the purchaser?

Yes, it is acceptable.

- 1.1.5 item 40 - Roundhole suction tube with 16 holes,diam. 5mm x 330mm Is it acceptable to offer Roundhole suction tube with holes, diam.5mm x 330mm (we do not have exact information about number of holes)

- 1.1.8 item 18 - Scalpel blade remover to remove the blades from handle without the risk of injury, 14cm Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

- 1.1.8 item 30 - Round bowl 145x65mm, 0,70l Is it acceptable to offer round bowl with 0,3 l deviation.

- 1.1.9 item 16 - Scalpel blade remover to remove the blades from handle without the risk of injury, 14cm.

Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

Yes, it is also acceptable.

- 1.1.9 item 19 - Converse needle holder with tungsten carbide inserts 0,4mm, 10,5cm with big sized rings.

Is it acceptable instrument with 2,5 cm deviation in length and 0,2mm deviation in tungsten carbide inserts?

Yes, it is also acceptable.

- 1.1.10 item 6 - Legueu bladder spatula, 105x35mm, 27cm Is it acceptable to offer spatula with 15 mm deviation in length?

Yes, it is also acceptable.

- 1.1.10 item 9 - Allis intestinal and tissue grasping forceps, 5x6 teeth, 22,5cm. Is it acceptable to offer spatula with 3 mm deviation in length?

Yes, it is also acceptable.

- 1.1.10 item 17 - Scalpel blade remover to remove the blades from handles without risk of injury, 14cm.

Is it acceptable to offer the set without this item? This item is not necessary for set functionality, its use value is very small, as the scalpel blade can be removed from the handle completely safely without the special remover.

Yes, it is also acceptable.

-1.1.10 item 26 - Potts-Smith dissecting forceps with tungsten carbide inserts, 0,4mm serration, 21cm, with dissector end, golden handles.

Is it acceptable to offer dissecting forceps with tungsten carbide inserts, length 200 mm, ring shaped dissector end diameter 2,5 mm, gold handle?

Yes, it is also acceptable.

- 1.1.10 item 40 - Randall kidney stone forceps, 20,5cm. Is it acceptable to offer instrument with 2 cm deviation in length?

Yes, it is also acceptable.

-1.1.10 item 54 - Intestinal and tissue grasping forceps Babcock 23.5cm. Is it acceptable to offer instrument with 15 mm deviation in length?

Yes, it is also acceptable.

- 1.1.10 item 69 - Intestinal clamp forceps Doyen 21cm straight, soft elastic. Is it acceptable to offer instrument with 2,5 cm deviation in length?

Yes, it is also acceptable.

- 1.1.10 item 70 - Intestinal clamp forceps Doyen 21cm curved, soft elastic Is it acceptable to offer instrument with 2,5 cm deviation in length?

Yes, it is also acceptable.

-1.1.10 item 71 - Stitch helper (retractor for wound closing), 200mm Is it acceptable to offer retractor for wound closing 1,5 cm longer?

Yes, it is also acceptable.

Question 4:

- 1.1.10 item 75 - Deaver retractor 30.5cm, 100x75mm. Is it acceptable to offer retractor with 1,5 cm deviation in length?

Yes, it is also acceptable.

-1.1.10 item 77 - Fritsch retractor 26cm/ 60x85mm, handle vertical to the blade. Is it acceptable to offer retractor with 2,5 cm deviation in length?

Yes, it is also acceptable.

Public Procurement Committee



Marija Stanisavljević