

## AMENDMENT No. 1 TO PROCUREMENT DOCUMENTS

Issued on 18th of February 2022

### PROCUREMENT OF MEDICAL EQUIPMENT FOR CLINICAL-HOSPITAL CENTRE “DR DRAGIŠA MIŠOVIĆ-DEDINJE”

IOP/59-2021/UHI

In accordance with the Clause 8. Part 1. Bidding Procedures, Section I. Instructions to Bidders, Contents of Bidding Documents, Amendment of Bidding Documents of the Procurement Documents, Public Investment Management Office, No. 11 Nemanjina street, Republic of Serbia, as the Purchaser, hereby notifies all persons concerned for **Procurement of medical equipment for Clinical-hospital centre “Dr Dragiša Mišović-Dedinje” IOP/59-2021/UHI**, that there has been an amendment made in the Procurement Documents.

The following provisions of Tender Documents are hereby replaced as follows:

**1. Tender Documents are changed in Technical Specification, Lot 1, Surgery, Line item no. 1.11, ID 1.11.1 - Anesthesia system with monitor, no. 16, which reads as follows:**

- Minimal range of flow trigger: 0.2 - 15 L/min

**It is replaced and reads as follows:**

- Minimal range of flow trigger: 0.2 - 10 L/min

In all other aspects, Tender documents for the **Procurement of medical equipment for Clinical-hospital centre “Dr Dragiša Mišović-Dedinje” IOP/59-2021/UHI**, remains unchanged.

This Amendment to Procurement Documents is a constituent part of the **Procurement of medical equipment for Clinical-hospital centre “Dr Dragiša Mišović-Dedinje” IOP/59-2021/UHI**, and will be posted on the Purchaser's web site: <http://www.obnova.gov.rs/english/public-procurement>

Updated version of technical specifications are given in excel file and is constituent part of this amendment.

In order to submit a responsive bid, bidders are expected to prepare their offers in accordance with this amendment.

Public Procurement Committee

