

| | | | | | |
|----------------|----------------------------|--|--|--|--|
| Bidder: | Biomedica MP d.o.o. | | | | |
| LOT11 | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| Line item No. | ID | Traženo | Ponudeno | QTY | |
|---------------|----|---|--|-----|---|
| 1 | | Cardio instruments | | 1 | |
| 1 | | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 2 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristic of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 2 | | Amputation instruments | | 1 | |
| 1 | | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 4 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristic of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 3 | | Anesthesia instruments | | 1 | |
| 1 | | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 8 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristic of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 4 | | AO Instruments | | 1 | |
| 1 | | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 7 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristic of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 4 | | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 Clamp, Atraumatic, acc. DeBakey, del 12.5cm | 10 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |

| | | | | | |
|----|--|---------|---|----|--|
| 34 | KOCHER Depressor Spoon, 220 mm (8 3/4"), spoon shaped, width: 28 mm, non-sterile, reusable | 77-9981 | Tablespoon | 1 | The requested specification is KOCHER depressor spoon, not Tablespoon |
| 5 | Electrophysiology instruments | | | 1 | |
| 3 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 10 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 4 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 | Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | 20 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |
| 6 | Addition to EP | | | 1 | |
| 1 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 2 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 7 | Femoropolieta instruments | | | 1 | |
| 5 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 4 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 8 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 | Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | 6 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |
| 8 | II B ICU | | | 1 | |
| 1 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 10 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 9 | Inv. Cardiology instruments | | | 1 | |

| | | | | | |
|----|--|---------|---|----|--|
| 1 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 | Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | 10 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |
| 3 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 10 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 10 | Cardiac surgery ICU | | | 1 | |
| 1 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 13 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 11 | Cardiac surgery set | | | 1 | |
| 20 | Castroviejo Micro Scissors, spring style, round handle, 60° angled micro fine blades, 4-1/4" / 11cm | 03-1044 | Scissors, Micro, Coronary, Castroviejo flat handle 9cm 60° edge 10mm | 1 | Requested length is 11cm, the offered is 9cm, which is significantly smaller then needed and specified |
| 21 | Castroviejo Micro Scissors, spring style, round handle, 90° angled micro fine blades, 4-1/8" / 10.5cm | 03-1046 | Scissors, Micro, Coronary, Castroviejo flat handle 9cm 90° edge 10mm | 1 | Requested length is 10.5cm, the offered is 9cm, which is significantly smaller then needed and specified |
| 22 | Castroviejo Micro Scissors, spring style, round handle, 125° angled micro fine blades, 4-1/8" / 10.5cm | 03-1048 | Scissors, Micro, Coronary, Castroviejo flat handle 9cm 125° edge 10mm | 1 | Requested length is 10.5cm, the offered is 9cm, which is significantly smaller then needed and specified. |
| 24 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 | Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | 20 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument |
| 12 | Carotide instruments | | | 1 | |
| 7 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18cm | 24-0351 | Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | 3 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no. 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate |
| 11 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 | Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | 10 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |

| | | | | | |
|-----------|--|---|--|----------|---|
| 13 | Cardiac surgery ward instruments | | | 1 | |
| 1 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | | 10 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |
| 14 | Urgent interventions Tmaponade instruments | | | 1 | |
| 9 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | | 10 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |
| 15 | Valvule cardiac surgery set | | | 1 | |
| 9 | Vascular haemostat, ring handle, curved jaws with sharp grasping serrations, 5" / 12.7 cm | 22-0939 Clamp, Atraumatic, acc. DeBakey, del. 12.5cm | | 5 | The requested instrument is not clamp, but vascular forceps-hemostat, mosquito type which is described with characteristic-jaws with sharp grasping serrations. The offered specification is not acceptable. The offered instrument is with jaws that are not tiny and sharp to reach around objects and hold the object securely in the tips of instrument. |
| 19 | KOCHER Depressor Spoon, 220 mm (8 3/4"), spoon shaped, width: 28 mm, non-sterile, reusable | 77-9981 Tablespoon | | 1 | The requested specification is KOCHER depressor spoon, not Tablespoon |
| 16 | Vasucelar ICU instruments | | | 1 | |
| 2 | Fell Needle Holder, 2 mm carbon-inlay jaws, gold ring handle, 7" / 18 cm | 24-0351 Holder, Needle, TC, Microvascular 18cm TC (M) rings gold plated | | 10 | Carbon-inlay value of the jaws is not stated in TSO, nor in catalogue. In public published catalogue on web, the purchaser found that instrument with catalogue no 24-0351 has 1.2 mm value which is characteristics of fine needle holder, but more rough is requested with specified 2mm. For our procedures needle holder with this characteristic is not appropriate. |